

Heartsine Technologies Ltd
 Canberra House
 203 Airport Road West
 Belfast
 Ireland BT3 9ED

Page 1 of 1

Certificate of Test

Electrical Safety for DEFIBRILLATOR

Date of Issue: 17th September 2010

SGS Reference: ELS126959/1/Amd 2/C/RG/09
 (This certificate replaces and supersedes ELS 126959/1/C/RG/09)

Details of Sample: Model: Heartsine Samaritan PAD, SAM300P
 Rated: Non rechargeable lithium battery 18V, 800mA Pad-Pak, 0 – 50°C,
 IP56, Defibrillation proof with type BF applied parts. 150/150/200J

Date of Receipt: 11th June 2009

Date of Test: 17th June 2009 to 22nd July 2009

Test Specification: Test for compliance with *EN 60601-2-4:2003 (Particular requirements for the safety of cardiac defibrillators) in conjunction with EN 60601-1:1990 +A1:93, A11:93, A12:93, A2:95 & A13:96 (Medical electrical equipment – General requirements for safety) and *ANSI/AAMI DF80 :2003 (excluding Biocompatibility, EMC, PEMS, Usability, Alarm systems & Equipment systems). IP not re-assessed.

Conclusion: The sample tested was found to comply with the Test Specification, based upon previous testing and assessment of the changes to the device.

Test Engineer:



M. W. Garfitt

Counter Signatory:



A. Pers

All enquires relating to this certificate must be directed to the Test Engineer

This document is issued by the Company subject to its General Conditions of Service available on request or accessible at www.sgs.com/terms_and_conditions.htm and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at www.sgs.com/terms_e-document.htm. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 28 days only.