



CERTIFICATE

Page 1 of 1

No. P03100705

Order No. 200311323

Applicant **LUXO AS**
Enebakkveien 117
Postboks 60 Manglerud
0612 Oslo, Norway

Manufacturer **LUXO AS**
Enebakkveien 117
Postboks 60 Manglerud
0612 Oslo, Norway

Factory **LUXO Industrier AS**
Postboks 54
2260 Kirkenær
Norway

Group 61 91000 **General fluorescent lamp luminaire, for work place**

Model/type **Circus Medical**

Data **230-240VAC, 50Hz, 400mA**

Other specification **Cl. I, Type B applied part.**

The above product is certified according to the following standard(s) Safety std.: **EN 60601-1 (1990) + A1 (1993) + A2 (1995) + A13 (1996) + Corrigenda (July 1994)**
MEDICAL ELECTRICAL EQUIPMENT - Part 1: General requirements for safety
 EMC std.: **EN 60601-1-2 (1993)**

Validity This certificate documents conformity with the standards shown, and also applies as license for use of Nemko's name and certification mark. The certificate and license is valid as long as the applicable conditions are complied with, and provided that any changes to the product are notified to Nemko for acceptance prior to implementation.
 New standards or amendments to the standards may imply that the product design must be updated and/or that re-testing and re-certification is necessary.

Additional information

Date of issue 21 May 2003

signature

Skule Moe

Certification Department

Nemko AS
 P.O. Box 73, Blindern
 N-0314 Oslo, Norway

Office address
 Gaustadalléen 30
 Oslo

Telephone
 +47 22 96 03 30
 Enterprise number:

Fax
 +47 22 96 05 50
 NO 974404532



Page 2 of 2

Annex to Licence

ENEC NO. 796
Order No. 200308249

Note: This Annex is an integral part of the LUM Licence and should only be read in connection with the Licence.

Licensee: (Manufacturer)	LUXO ASA Enebakkveien 117 Postboks 60, Manglerud 0612 OSLO
Trade mark	LUXO
Model/type:	Circus
The product is certified according to the following standards	EN 60 598-2-4:97 EN 60 598-1:2000 +A11:2000 + A12:2002
Rated voltage	230-240V
Rated wattage and number of lamps	1x 22W
Lamp type and attachment:	Fluorecent lamp, 2GX13 lamp holder
Degree of protection: (IPXX)	IP 20
Mains connection	Plug
Classification regarding supporting material	F-symbol, normally flammable surface
Protection against electric shock: (Class I, II, III)	Class I
Additional Information	

Summary of limitations which has been considered.
(Including the countries adhering to the scheme, for which the luminaire is not certified)

Superseded licence:

Note: This Annex is an integral part of the LUM Licence and should only be read in connection with the Licence.

Nemko AS P.O. Box 73, Blindern N-0314 Oslo, Norway	Office address Gaustadalléen 30 Oslo	Telephone +47 22 96 03 30 Enterprise number:	Fax +47 22 96 05 50 NO 974404532
--	--	--	--