

Heartsine Technologies Ltd
 Canberra House
 203 Airport Road West
 Belfast
 Ireland
 BT3 9ED

Page 1 of 1

Certificate of Test

Electrical Safety

For

Portable Defibrillator Unit

Date of Issue: 18th August 2006

SGS Reference: ELS 102384/1/C/RG/06

Details of Sample: Model : Samaritan Pad "SAMPAD" P/N - SAM 300
 Serial No : 04-0001356

Date of Receipt: 17th July 2006

Date of Test: 20th July 2006 to 11th August 2006

Test Specification: Test for compliance with IP56 as described in BS EN 60529:1992 incorporating Amendments 1 & 2 (Degrees of protection provided by enclosures (IP code))


Conclusion: The sample tested was found to comply with IP5X (category 1) and IPX6. and therefore complies with IP56 as required by the standard.

This certificate is issued by the Company subject to its General Conditions of Service available upon request and accessible at www.sgs.com. Attention is drawn to the limitations of liability, indemnification, jurisdictional, representative restrictions of test samples and sample retention policies defined therein.

This certificate refers only to the sample submitted for test.

This certificate shall not be reproduced except in full without the written approval of SGS United Kingdom Ltd.

Authorised Signatory: R. N. Gough



Counter Signatory : S. Wilkinson



All enquires relating to this certificate must be directed to the Authorised Signatory

QUAL 08 Rev 5